

## Konformitätserklärung Declaration of Conformity

**Erzeugnis**  
*Product*

Druckmessgerät  
*pressure measurement instrument*

**Baureihe/Typ**  
*Type*

Kapselfederanometer KP 100  
*pressure gauge with capsule element*

**Anzeigebereiche**  
*Ranges*

0 – 25 mbar  
0 – 40 mbar  
0 – 60 mbar  
0 – 100 mbar  
0 – 160 mbar  
0 – 250 mbar  
0 – 400 mbar

**Richtlinien/Normen**  
*Directives/Standards*

DIN EN 837-3: 02/1997 (Typ KP 100 ...)


Wir erklären, dass die oben genannten Produkte die Anforderungen der aufgeführten Richtlinien und Normen erfüllen. Diese Manometer fallen nicht in den Geltungsbereich der Druckgeräterichtlinie (2014/68/EU) und dürfen keine CE-Kennzeichnung tragen.

*We declare that the above mentioned products fulfil the requirements of the list directives and standards. These pressure gauges do not fall within the scope of the Pressure Equipment Directive (2014/68/EU) and must not have a CE marking.*

Wallenhorst, 29.04.2020

HEINZ MARCHEL  
GMBH & CO. KG  
Ringstr. 3  
D-49134 Wallenhorst

  
.....  
( ppa. Helmut Sienkamp )

  
.....  
( i. A. Marcus Menzel )  
(techn. Leitung / Manager Technology)

KON KP100/25-400 mbar 29.04.20 DE/GB